

Comparison of the electrochemical sensitivity of modified sensors based on screen-printed electrodes for real-time detection of Doxorubicin

Iva Dimitrievska*, Perica Paunović, Anita Grozdanov

Ss. Cyril and Methodius University in Skopje, Faculty of Technology and Metallurgy, Rugjer Boshković 16,
1000 Skopje, North Macedonia *iva@tmf.ukim.edu.mk

Last decade, an intensive research effort has been performed in the field of analytical electrochemistry seeking designs of electrochemical biosensors capable of providing better analytical characteristics in terms of sensitivity, selectivity, reliability, ease of fabrication and use, and low cost, especially for pharmaceutical drugs' monitoring. The incorporation of nanomaterials into composite electrode matrices presents a useful strategy for the preparation of electrochemical biosensors with improved analytical performance. These devices exhibit the characteristics of the used nanomaterials and the advantages of composite electrodes, such as low background currents, great versatility due to the possibility of incorporation of different substances into the bulk of the electrode matrix, and easy surface regeneration.

Doxorubicin hydrochloride (DOX), a potent anticancer drug, is widely used in chemotherapy due to its high effectiveness against a broad spectrum of malignancies, including breast cancer, lymphomas, leukemia, and sarcomas. Doxorubicin's anticancer and antitumor main mechanism of action is related to the inhibition of DNA replication and transcription processes through intercalation between base pairs, thereby leading to cell death¹.

The research in our work has been primarily focused on the development of screen-printed electrode (SPE)-based sensors and their application as electrochemical platforms for drug determination and monitoring, with a specific emphasis on their suitability for surface modification. A commercial screen-printed graphene electrode was used as the electrochemical sensing component, which was subsequently modified with polymers. The polymeric modification, achieved by coating onto electrode surfaces, offers flexibility and introduces diverse functional groups that contribute to enhancing the reactivity of the materials². Polyvinylidene fluoride and chitosan were used for polymer modification. All studied electrodes were tested using a solution of 0.002 mol/L DOX dissolved in 0.1 M phosphate-buffered saline with a pH of 6.8. Cyclic voltammetry was used as an electrochemical characterization technique, to gather data regarding the electrochemical activity of all developed electrodes. The physical characterization of the electrodes was done using scanning electron microscopy (SEM). The changes in the electrolyte during the electrochemical measurements were followed through ultraviolet-visible (UV-Vis) spectroscopy. All modified electrodes demonstrate a favourable electrochemical response to DOX and exhibit higher electrical conductivity compared to commercial ones. The obtained results from the characterization indicate that polymer-modified polyaniline electrodes exhibit excellent electrochemical conductivity and demonstrate the best overall electrochemical performance.

References:

1. C. Carvalho, RX. Santos, S. Cardoso, S. Correia, PJ. Oliveira, MS. Santos, PI. Moreira, Doxorubicin: the good, the bad and the ugly effect. *Curr Med Chem.* 2009;16(25):3267-85 <https://doi.org/10.2174/092986709788803312>
2. J. Soleymani, M. Hasanzadeh, M. Eskandani, M. Khoubnasabjafari, N. Shadjou, A. Jouyban, Electrochemical sensing of doxorubicin in unprocessed whole blood, cell lysate, and human plasma samples using thin film of poly-arginine modified glassy carbon electrode, *Materials Science and Engineering: C*, Volume 77, 2017, Pages 790-802, ISSN 0928-4931, <https://doi.org/10.1016/j.msec.2017.03.257>